FYS-KJM 4740

MR-teori og medisinsk diagnostikk

Kap 6
k-space acceleration

Atle Bjørnerud, Rikshospitalet
atle.bjornerud@fys.uio.no
975 39 499

Fast Spin Echo (FSE)

Acquisition of multiple k-lines per TR

FSE



FSE



FSE – effective echo time:



FSE – echo modulation:



Signal (as fn of k-space line) is modulated by T2-relaxation with point-spread function given by:

$$S(P) = \int_{-k_{max}/2}^{k_{max}/2} P(k_y)\exp(-jk_y)dk_y$$

P(ETL)=exp(-ETL·ES/T2)

FSE – echo modulation:



Signal (as fn of k-space line) is modulated by T2-relaxation with point-spread function given by:

$$S(P) = \int_{-k_{max}/2}^{k_{max}/2} P(k_y)\exp(-jk_y)dk_y$$ P(ETL)=exp(-ETL·ES/T2)

FSE – echo modulation:

$$S(P) = \int_{-k_{max}/2}^{k_{max}/2} P(k_y)\exp(-jk_y)dk_y$$

⇩ Exp T2-decay =>Lorenzian kernel

$$S(y,T2) = \frac{1/W}{1+j2y/W}$$ W=δy·2/π (ETL · ES/T2)

Δy= pixel dim in phase enc-dir



FSE – Different profile orders:

linear

centric



Segmented FSE:
K-space 'raster': minimizing k-space discontinuities

- group together lines with equal attenuation
- minimum step size between adjacent blocks of equally attenuated lines.



FSE :





Echo Planar Imaging (EPI)

Similar to FSE but using GRE in stead of SE …



SE or GRE 'preparation'





Eddy currents:

$$B_z \cong B_0\left\{1+\frac{G_z^2 z^2}{2B_0^2}\right\}+G_x x$$



Spiral imaging

Trajectory for mth segment:

$\mathbf{k}(t) = A\varphi(t)\exp(j\varphi(t) + jm\varphi_0)$; $\varphi(t) = t\pi/N$