

BAKERType : J02 Reading : 0.01 mm
Tested in accordance to : DIN-878

Range : 10 mm

Graduation : 0 - 100

QA/ 29 / 00.00
Gauge Serial No : B 1232
Certificate No. : 6131

	0	10	20	30	40	50	60	70	80	90
0	0.000	0.000	-0.002	-	-	-	0.002	0.001	-0.001	-
1	-	-	0.001	0.002	0.001	-	-	-	0.001	0.001
2	0.002	-	-	-	-0.001	0.001	0.002	-	-	-
3	0.001	0.000	0.002	-	-	-	0.003	0.003	0.001	-
4	-	-	0.000	0.002	0.003	-	-	-	-0.001	0.000
5	0.000	-	-	-	-0.001	0.001	0.002	-	-	-
6	-0.001	0.000	0.000	-	-	-	0.002	0.001	0.002	-
7	-	-	0.000	0.000	0.001	-	-	-	-0.001	-0.001
8	0.001	-	-	-	0.003	0.004	0.004	-	-	-
9	0.003	0.001	0.003	-	-	-	0.005	0.003	0.004	-
10	-	-	-	-	-	-	-	-	-	-

Calibration Equipment used is traceable to National Length Standards vide NPL certificate no. 1038/07/664/98.

Errors in units of : mm

Uncertainty of measurement : 0.001 mm

A

Q.A.

**AN ISO9001 COMPANY**