

BAKER

CERTIFICATE OF CALIBRATION

QW/29/00.00

Type : J02 Reading : 0.01 mm

Range : 10 mm

Graduation : 0 - 100

Gauge Serial No. : B 1356

Tested in accordance to : DIN-878

Certificate No. : 6132

	0	10	20	30	40	50	60	70	80	90
0	0.000	-0.001	-0.002	-	-0.001	-	0.000	0.000	0.002	0.003
1	-	-	-0.002	-	0.000	-	-	-	0.000	-
2	0.002	-	-	-	0.000	0.000	0.000	-	-	-
3	0.002	0.002	0.001	-	-	-	0.001	0.002	0.003	0.003
4	-	-	0.000	0.000	0.001	-	-	-	0.002	0.004
5	0.004	-	-	-	0.000	0.001	0.002	-	-	-
6	0.003	0.003	0.001	-	-	-	0.000	0.003	0.004	-
7	-	-	0.001	0.000	0.001	-	-	-	0.002	0.003
8	0.004	-	-	-	0.002	0.000	0.003	-	-	-
9	0.006	0.001	0.002	-	-	-	0.006	0.003	0.003	-
10	-	-	-	-	-	-	-	-	-	-

Calibration Equipment used is traceable to National Length Standards vide NPL certificate no. 1039A/07/664/98.

Errors in units of : mm

Uncertainty of measurement : 0.001 mm

AN ISO9001 COMPANY



Q.A.