

BAKERType : J02 Reading : 0.01 mm
Tested in accordance to : DIN-878**CERTIFICATE OF CALIBRATION**

Range : 10 mm

Graduation : 0 - 100

QA/29/00.00

Gauge Serial No : B 1605


Certificate No. : 7051

	0	10	20	30	40	50	60	70	80	90
0	0.000	0.001	-0.001	-	-	-	-0.004	-0.003	-0.003	-
1	-	-	0.001	-0.003	-0.002	-	-	-	-0.002	-0.003
2	0.000	-	-	-	-0.005	-0.005	-0.005	-	-	-
3	-0.003	-0.002	-0.003	-	-	-	-0.007	-0.004	-0.004	-
4	-	-	-0.003	-0.005	-0.006	-	-	-	-0.005	-0.003
5	-0.004	-	-	-	-0.007	-0.008	-0.009	-	-	-
6	-0.007	-0.005	-0.006	-	-	-	-0.010	-0.008	-0.005	-
7	-	-	-0.005	-0.005	-0.005	-	-	-	-0.003	0.000
8	-0.001	-	-	-	-0.005	-0.004	-0.004	-	-	-
9	0.001	0.000	0.000	-	-	-	-0.004	-0.004	-0.002	-
10	-	-	-	-	-	-	-	-	-	-

Calibration Equipment used is traceable to National length Standards vide NPL certificate no. 1038/07/664/98.

Errors in units of : mm

Uncertainty of measurement : 0.001 mm

 Q.A.
**AN ISO9001 COMPANY**