

BAKER**CERTIFICATE OF CALIBRATION**

QA/29/00.00

Type : J02

Reading : 0.01 mm

Range : 10 mm

Graduation : 0 - 100

Gauge Serial No. : C 3767

Tested in accordance to : DIN-878

Certificate No. : 6972

	0	10	20	30	40	50	60	70	80	90
0	0.000	-0.002	-0.004	-	-	-	0.001	0.000	-0.002	-
1	-	-	0.000	-0.002	-0.003	-	-	-	0.000	0.002
2	-0.001	-	-	-	-0.001	-0.001	-0.002	-	-	-
3	0.000	-0.004	-0.007	-	-	-	-0.001	0.000	-0.002	-
4	-	-	0.000	-0.001	-0.001	-	-	-	-0.001	-0.001
5	0.002	-	-	-	-0.004	-0.001	0.000	-	-	-
6	0.002	0.001	-0.001	-	-	-	0.000	0.001	0.001	-
7	-	-	0.002	0.000	0.000	-	-	-	0.000	0.002
8	0.004	-	-	-	0.000	0.001	0.005	-	-	-
9	0.002	0.003	0.003	-	-	-	0.002	0.002	0.003	-
10	-									

Calibration Equipment used is traceable to National length Standards vide NPL certificate no. 1038/07/664/98.

Errors in units of : mm

Uncertainty of measurement : 0.001 mm

Q.A.

**AN ISO9001 COMPANY**