

BAKER**CERTIFICATE OF CALIBRATION**

QA/29/00.00

Type : J02 Reading : 0.01 mm
Tested in accordance to : DIN-878

Range : 10 mm

Graduation : 0 - 100

Gauge Serial No : C 4352
Certificate No. : 6980

| | 0 | 10 | 20 | 30 | 40 | 50 | 60 | 70 | 80 | 90 |
|----|--------|--------|--------|--------|--------|--------|--------|--------|--------|--------|
| 0 | 0.000 | -0.003 | -0.004 | - | - | - | -0.003 | -0.005 | -0.006 | - |
| 1 | - | - | -0.003 | -0.003 | -0.006 | - | - | - | -0.002 | -0.001 |
| 2 | -0.004 | - | - | - | -0.005 | -0.003 | -0.005 | - | - | - |
| 3 | 0.000 | 0.000 | -0.003 | - | - | - | -0.001 | -0.001 | -0.001 | - |
| 4 | - | - | -0.001 | -0.003 | -0.004 | - | - | - | 0.002 | 0.002 |
| 5 | 0.003 | - | - | - | -0.002 | 0.001 | 0.002 | - | - | - |
| 6 | 0.003 | 0.005 | 0.002 | - | - | - | 0.000 | 0.003 | 0.002 | - |
| 7 | - | - | 0.001 | 0.000 | 0.000 | - | - | - | 0.003 | 0.003 |
| 8 | 0.004 | - | - | - | -0.001 | 0.000 | 0.002 | - | - | - |
| 9 | 0.003 | 0.004 | 0.004 | - | - | - | 0.000 | 0.001 | 0.003 | - |
| 10 | - | - | - | - | - | - | - | - | - | - |

Calibration Equipment used is traceable to National length Standards vide NPL certificate no. 1038/07/664/98.

Errors in units of : mm

Uncertainty of measurement : 0.001 mm

16 Q.A.

**AN ISO9001 COMPANY**